

## DECLARATION OF CONFORMITY

No.HATCE0406A

### ■ WE THE MANUFACTURER and

- Name **HORIBA Advanced Techno, Co., Ltd.**  
- Address 2 Miyanohigashi-cho, Kisshoin, Minami-ku, Kyoto, 601-8551, Japan

### ■ THE AUTHORISED REPRESENTATIVE IN EU

- Name HORIBA Europe GmbH  
- Address Hans-Mess-Str.6, D-61440 Oberursel, Germany

### ■ DECLARE THAT THE PRODUCT

- Designation pH Meter, COND Meter  
- Model name F-72G, F-73G, F-74G, DS-72G

### ■ CONFORMS TO DIRECTIVES AND STANDARDS

- Directive(s)	The EMC Directive	2014/30/EU
	The Low Voltage Directive	2014/35/EU
	The RoHS Directive	2011/65/EU and (EU)2015/863
- Standard(s)	The EMC Directive	EN 61326-1:2013
	- Emission	Class B
	- Immunity	Basic electromagnetic environment
	The Low Voltage Directive	EN 61010-1:2010/A1:2019
	The RoHS Directive	EN IEC 63000:2018
	- Product Category	9. Monitoring and control instruments


#### ■ Manufacturer

10 Jun. 2021  
Date of issue

  
Signature

**Yoshiyasu Ito**  
General Manager  
Quality Assurance Office

#### ■ Authorised Representative in EU

  
Signature

**Stefan Grabert**  
EQS Quality Manager